

VISHNU DENTAL COLLEGE :: BHIMAVARAM
Documentation Sheet to Monitor CSSD Sterilization Process
Sterilizer No. _____

Date: _____

Sterilization Process: Steam
Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VHICSA110	VHICSR110	VHICSA110	VHICSR110
Date of production	26/04/23	27/04/23	27/04/23	28/04/23
Expiry Date	30/05/23	30/05/23	30/05/23	30/05/23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20PM	1:20PM	5:20PM	1:25PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICSA110	VHICSR110	VHICSA110	VHICSA110
Date of production	28/04/23	29/04/23	29/04/23	01/05/23
Expiry Date	30/05/23	30/05/23	30/05/23	30/05/23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30PM	1:15PM	5:20PM	2:00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)


BDS-test indicator strip	Test OK?		Responsible User
	☐ Yes	☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH1C5A110	VH1C5B110	VH1C5A110	VH1C5B110
Date of production	02/05/23	03/05/23	04/05/23	05/05/23
Expiry Date	30/05/23	30/05/23	30/05/23	30/05/23
Stick in Indicator strip				
Program	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle
Temperature Starting Time	121°C 2.10PM	121°C 2.30PM	121°C 2.20PM	121°C 2.15PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VH1C5A110	VH1C5B110	VH1C5A110	VH1C5A110
Date of production	6/05/23	8/05/23	9/05/23	10/05/23
Expiry Date	30/05/23	30/05/23	30/05/23	30/05/23
Stick in Indicator strip				
Program	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle
Temperature Starting Time	121 °C 2.30PM	121°C 2.20PM	121°C 2.30PM	121 °C 2.20PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



Ster

Ster

Bow

is

10/05/23

STEAM	LOT 201044	121°C / 10 min.	
		134°C / 3,5 min.	

Batch Monitoring*

User, sterilizer & batch number	VH105A110	VH105B110	VH105A110	VH105B110
Date of production	10-05-23	11-05-23	11-05-23	12-05-23
Expiry Date	30-05-23	30-05-23	30-05-23	30-05-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	1.00PT	2.10PT	2.15PT	2.20PT
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VH105A110	VH105A110	VH105B110	VH105A110
Date of production	12-05-23	15-05-23	16-05-23	17-05-23
Expiry Date	30-05-23	30-05-23	30-05-23	30-05-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	2.30PT	2.00PT	2.20PT	2.30PT
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi



355

18/05/23

Ste
Ste
Bo

STEAM

101 201044

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	18-05-23	19-05-23	20-05-23	22-05-23
Expiry Date	30-05-23	30-05-23	30-05-23	30-05-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	2:00 PM	2:05 PM	2:30 PM	2:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. Jayanthi	10. Jayanthi	10. Jayanthi	10. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	23-05-23	24-05-23	25-05-23	26-05-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	2:00 PM	2:30 PM	2:30 PM	2:15 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. Jayanthi	10. Jayanthi	10. Jayanthi	10. Jayanthi

65(A) 3-06-23

SAL Steri-test BOWIE&DICK TEST PACK



Sterilize

Steriliza

Bowie-D

06-23

pr

STEAM

LOT 201044

121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTR110	VHICTA110	VHICTR110
Date of production	27-05-23	29-05-23	30-05-23	31-05-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Current cycle	Current cycle	Current cycle	Current cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	2:00PM	2:20PM	2:30PM	2:40PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTR110	VHICTA110	VHICTR110
Date of production	1-06-23	2-06-23	3-06-23	5-06-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Current cycle	Current cycle	Current cycle	Current cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20PM	5:30PM	5:40PM	5:40PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK



Sterilizer No

Sterilization

Bowie-Dick

BDS-

STEAM

LOT 190414

134 °C / 3,5 min.

-23

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	6-06-23	7-06-23	8-06-23	9-06-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Excurrent cycle	Excurrent cycle	Excurrent cycle	Excurrent cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:10 PM	5:20 PM	5:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTA110	VHICTB110	VHICTA110
Date of production	10-06-23	12-06-23	13-06-23	14-06-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Excurrent cycle	Excurrent cycle	Excurrent cycle	Excurrent cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:30 PM	5:00 PM	5:30 PM	5:45 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☒ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTA110	VHICTA110	VHICTA110
Date of production	15-06-23	15-06-23	16-06-23	17-06-23
Expiry Date	30-06-23	30-06-23	30-06-23	30-06-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30 PM	5:20 PM	5:30 PM	6:00 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTA110	VHICTA110	VHICTA110
Date of production	19-06-23	20-06-23	20-06-23	21-06-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30 PM	1:10 PM	5:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?		Responsible User
	☐ Yes	☐ No	

Batch Monitoring*

User, sterilizer & batch number	VHICJA110	VHICIB110	VHICJA110	VHICIB110
Date of production	21-06-23	22-06-23	24-06-23	25-06-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20PM	5:30PM	5:40PM	5:45PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICJA110	VHICIB110	VHICJA110	VHICIB110
Date of production	26-06-23	27-06-23	28-06-23	29-06-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:30PM	5:20PM	5:30PM	5:40PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



Access

e: 3/07/23

STEAM

LOT 201044

121°C / 10 min.

34°C / 3,5 min.

the

Batch Monitoring*

User, sterilizer & batch number	VHKTANO	VHKTANO	VHKTANO	VHKTANO
Date of production	30-06-23	01-07-23	01-07-23	3-07-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature Starting Time	121°C 5:20pm	121°C 5:30pm	121°C 5:40pm	121°C 5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	ic. Jayanthi	ic. Jayanthi	ic. Jayanthi	ic. Jayanthi

User, sterilizer & batch number	VHKTANO	VHKTANO	VHKTANO	VHKTANO
Date of production	4-07-23	5-07-23	6-07-23	7-07-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature Starting Time	121°C 5:30pm	121°C 5:20pm	121°C 5:30pm	121°C 5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	ic. Jayanthi	ic. Jayanthi	ic. Jayanthi	ic. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



AM

tion Process

Date: 8/7/23

STEAM

LOT 201044

121°C / 10 min.
134°C / 3,5 min.

User

Jayanthi

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	8-07-23	10-07-23	11-07-23	12-07-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:00pm	5:20pm	5:10pm	5:20pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	13-07-23	14-07-23	15-07-23	17-07-23
Expiry Date	30-07-23	30-07-23	30-07-23	30-07-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30pm	5:30pm	5:30pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?		Responsible User
	☐ Yes	☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH103A110	VH103B110	VH103A110	VH103B110
Date of production	18-07-23	19-07-23	20-07-23	21-07-23
Expiry Date	30-07-23	30-07-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Current cycle	Current cycle	Current cycle	Current cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20pm	5:30pm	5:20pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VH103A110	VH103B110	VH103A110	VH103B110
Date of production	22-07-23	24-07-23	25-07-23	26-07-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Current cycle	Current cycle	Current cycle	Current cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:00pm	5:10pm	5:20pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



Access

e: 2/08/23

STEAM

LOT 190414

134 °C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTR110	VHICTA110	VHICTR110
Date of production	27-07-23	28-07-23	29-07-23	31-07-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature Starting Time	121 °C 5:10PM	121 °C 5:20PM	121 °C 5:20PM	121 °C 5:30PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

User, sterilizer & batch number	VHICTA110	VHICTR110	VHICTA110	VHICTR110
Date of production	1-08-23	2-08-23	3-08-23	4-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature Starting Time	121 °C 5:20PM	121 °C 5:20PM	121 °C 5:30PM	121 °C 5:30PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE-DICK TEST PACK



De
Sterilizer No.
Sterilization
Bowie-Dick-

123

BDS-test	STEAM	LOT 201044	121°C / 10 min. 134°C / 3,5 min.
----------	--------------	------------	-------------------------------------

Batch Monitoring*

User, sterilizer & batch number	VHKTAL10	VHKTBI10	VHKTAL10	VHKTBI10
Date of production	4-08-23	5-08-23	5-08-23	7-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature Starting Time	121°C 5.40PM	121°C 3.30PM	121°C 5.40PM	121°C 2.00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHKTAL10	VHKTBI10	VHKTAL10	VHKTAL10
Date of production	7-08-23	8-08-23	8-08-23	9-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature Starting Time	121°C 5.30PM	121°C 1.30PM	121°C 5.40PM	121°C 5.30PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?		Responsible User
	☐ Yes	☐ No	

Batch Monitoring*

User, sterilizer & batch number	VHICTB110	VHICTA110	VHICTB110	VHICTA110
Date of production	10-08-23	11-08-23	12-08-23	12-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Exhaust cycle		Exhaust cycle	
Temperature	121°C	5:40PM	121°C	5:40PM
Starting Time				
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanthi		K. Jayanthi	

User, sterilizer & batch number	VHICTB110	VHICTA110	VHICTA110	VHICTB110
Date of production	14-08-23	14-08-23	16-08-23	16-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Exhaust cycle		Exhaust cycle	
Temperature	121°C	1:20PM	121°C	5:30PM
Starting Time				
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanthi		K. Jayanthi	

SAL Steri-test BOWIE&DICK



Sterilize
Steriliza
Bowie-D

123

BI	STEAM	LOT 201044	121°C / 10 min. 134°C / 3,5 min.
----	--------------	------------	-------------------------------------

Batch Monitoring*

User, sterilizer & batch number	VH KTA 110	VH KTB 110	VH KTA 110	VH KTB 110
Date of production	17-08-23	17-08-23	18-08-23	19-08-23
Expiry Date	30-08-23	30-08-23	30-08-23	30-08-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	1:15 PM	5:30 PM	5:20 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VH KTA 110	VH KTB 110	VH KTA 110	VH KTB 110
Date of production	21-08-23	22-08-23	23-08-23	23-08-23
Expiry Date	31-08-23	31-08-23	31-08-23	31-08-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20 PM	5:30 PM	5:20 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



Steril

Steril

Bow



23/08/23

STEAM

LOT 201044

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	24-08-23	25-08-23	26-08-23	26-08-23
Expiry Date	30-09-23	30-09-23	30-09-23	30-09-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20PM	5:30PM	1:10PM	5:20PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTA110
Date of production	28-08-23	28-08-23	29-08-23	30-08-23
Expiry Date	30-09-23	30-09-23	30-09-23	30-09-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	1:15PM	5:30PM	5:30PM	5:40PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	1/09/23	4/09/23	5/09/23	6/09/23
Expiry Date	30/09/23	30/09/23	30/09/23	30/09/23
Stick in Indicator strip				
Program	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle
Temperature Starting Time	121 °C 5:30PM	121 °C 5:30PM	121 °C 5:40PM	121 °C 5:10PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	7/09/23	8/09/23	11/09/23	12/09/23
Expiry Date	30/09/23	30/09/23	30/09/23	30/09/23
Stick in Indicator strip				
Program	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle	Exhaust Cycle
Temperature Starting Time	121 °C 5:20PM	121 °C 5:20PM	121 °C 5:30PM	121 °C 5:20PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



123

Doc
Sterilizer No. .
Sterilization P
Bowie-Dick-Si

BDS-test	STEAM	LOT 201044	121°C / 10 min. 134°C / 3,5 min.
----------	--------------	------------	-------------------------------------

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	13-09-23	14-09-23	15-09-23	16-09-23
Expiry Date	30-09-23	30-09-23	30-09-23	30-09-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:15 PM	5:20 PM	5:10 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	19/09/23	20/09/23	21/09/23	22/09/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	6:00 PM	5:40 PM	5:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST



ISS

15/09/23

Ste

Ste

Bo

STEAM

LOT 190414


134 °C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTA110	VHICTB110	VHICTA110
Date of production	23/09/23	25/09/23	26/09/23	27/09/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20 PM	5:30 PM	5:20 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	28/09/23	29/09/23	30/09/23	31/09/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:40 PM	5:40 PM	5:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST



DC DENTAL COLLEGE

1/10/23

Sterili

Sterili

Bowie

STEAM

LOT 201044

121°C / 10 min.


134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	3/10/23	4/10/23	5/10/23	6/10/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30PM	5:20PM	5:30PM	5:40PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	7/10/23	9/10/23	10/10/23	11/10/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:00PM	5:30PM	5:40PM	5:50PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer
Sterilization
Bowie-D

10/23
2/23
B)

EO	STEAM	LOT 201044	121°C / 10 min. 134°C / 3,5 min.
----	--------------	------------	-------------------------------------

Batch Monitoring*

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	12/10/23	13/10/23	14/10/23	16/10/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:40PM	5:30PM	5:30PM	5:20PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	IC Jayanthi	IC Jayanthi	IC Jayanthi	IC Jayanthi

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	17/10/23	18/10/23	19/10/23	20/10/23
Expiry Date	30/10/23	30/10/23	30/10/23	30/10/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:15PM	5:15PM	5:40PM	5:00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	IC Jayanthi	IC Jayanthi	IC Jayanthi	IC Jayanthi

SAL Steri-test BOWIE & DICK TEST PACK



Sterilizer

Sterilizat

Bowie-Di



10/23
01/23
15)

BD

STEAM

LOT 190414

134 °C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH105A110		VH105B110		VH105A110		VH105B110	
Date of production	25/10/23		25/10/23		26/10/23		27/10/23	
Expiry Date	30/11/23		30/11/23		30/11/23		30/11/23	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	12:00 PM		5:00 PM		5:30 PM		5:40 PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	V. S. J. S.		V. S. J. S.		V. S. J. S.		V. S. J. S.	

User, sterilizer & batch number	VH105A110		VH105B110		VH105A110		VH105B110	
Date of production	28/10/23		28/10/23		30/10/23		31/10/23	
Expiry Date	30/11/23		30/11/23		30/11/23		30/11/23	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	5:30 PM		5:40 PM		5:20 PM		5:30 PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	V. S. J. S.		V. S. J. S.		V. S. J. S.		V. S. J. S.	

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer

Sterilizat

Bowie-D

BD

23

STEAM

LOT 190414


134 °C / 3,5 min.

Batch M

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	01-11-23	02-11-23	3-11-23	4-11-23
Expiry Date	30-11-23	30-11-23	30-11-23	30-11-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:00pm	5:30pm	5:20pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Japalle	K. Japalle	K. Japalle	K. Japalle

User, sterilizer & batch number	VHICTA110	VHICTB110	VHICTA110	VHICTB110
Date of production	6-11-23	7-11-23	8-11-23	9-11-23
Expiry Date	30-11-23	30-11-23	30-11-23	30-11-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20pm	5:20pm	5:00pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Japalle	K. Japalle	K. Japalle	K. Japalle

SAL Steri-test BOWIE & DICK TEST



23

Sterilizer No

Sterilization

Bowie-Dick

BDS-10

STEAM

LOT 230205

121°C / 10 min.

134°C / 3.5 min.

Batch Monitoring*

User, sterilizer & batch number	VH 105 A110	VH 105 B110	VH 105 A110	VH 105 B110
Date of production	10-11-23	11-11-23	13-11-23	14-11-23
Expiry Date	30-11-23	1-12-23	1-12-23	1-12-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20 PM	5:30 PM	5:20 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. J. Jagtap	U. J. Jagtap	U. J. Jagtap	U. J. Jagtap

User, sterilizer & batch number	VH 105 A110	VH 105 B110	VH 105 A110	VH 105 B110
Date of production	15-11-23	16-11-23	17-11-23	18-11-23
Expiry Date	5-12-23	5-12-23	5-12-23	5-12-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20 PM	5:15 PM	5:20 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. J. Jagtap	U. J. Jagtap	U. J. Jagtap	U. J. Jagtap

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH115B110	VH115B110	VH115B110	VH115B110
Date of production	18/11/23	20/11/23	20/11/23	21/11/23
Expiry Date	5/12/23	5/12/23	5/12/23	31/12/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	2.00pm	1.00pm	5.30pm	1.20pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. Jayanthi	10. Jayanthi	10. Jayanthi	10. Jayanthi

User, sterilizer & batch number	VH115B110	VH115B110	VH115B110	VH115B110
Date of production	21/11/23	22/11/23	22/11/23	23/11/23
Expiry Date	11/12/23	11/12/23	11/12/23	11/12/23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5.20pm	1.40pm	5.30pm	1.45pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. Jayanthi	10. Jayanthi	10. Jayanthi	10. Jayanthi


SAL Steri-test BOWIE & DICK TEST PACK



Sterilizer No

Sterilization

Bowie-Dick



BDS-te

STEAM

LOT 230205

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHICSA110	VHICSA110	VHICSA110	VHICSA110
Date of production	24-11-23	25-11-23	25-11-23	27-11-23
Expiry Date	15-12-23	15-12-23	15-12-23	15-12-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:00 PM	1:00 PM	5:10 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. J. J. J.	10. J. J. J.	10. J. J. J.	10. J. J. J.

User, sterilizer & batch number	VHICSA110	VHICSA110	VHICSA110	VHICSA110
Date of production	28-11-23	29-11-23	30-11-23	7-12-23
Expiry Date	15-12-23	15-12-23	15-12-23	30-12-23
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20 PM	5:30 PM	5:40 PM	5:20 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10. J. J. J.	10. J. J. J.	10. J. J. J.	10. J. J. J.

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☒ Yes ☐ No	<i>10. Jayanthi</i>

Batch Monitoring*

User, sterilizer & batch number	<u>VH10TA110</u>	<u>VH10TB110</u>	<u>VH10TA110</u>	<u>VH10TB110</u>
Date of production	8-12-23	8-12-23	9-12-23	9-12-23
Expiry Date	30-12-23	30-12-23	30-12-23	30-12-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	1.00 PM	5.20 PM	1.15 PM	5.30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>

User, sterilizer & batch number	<u>VH10TA110</u>	<u>VH10TA110</u>	<u>VH10TA110</u>	<u>VH10TB110</u>
Date of production	11-12-23	12-12-23	13-12-23	14-12-23
Expiry Date	30-12-23	30-12-23	31-12-23	31-12-23
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5.30 PM	5.40 PM	5.30 PM	4.00 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>	<i>10. Jayanthi</i>

SAL Steri-test BOWIE&DICK TEST PA



AGE

Docu

Sterilizer No.

Sterilization Pro

Bowie-Dick-Sim

BDS-test ind

STEAM

LOT 230205

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring

User, sterilizer & batch number	VH11TA110	VH11TB110	VH11TB110	VH11TB110
Date of production	14-12-23	15-12-23	15-12-23	16-12-23
Expiry Date	31-12-23	5-01-24	5-01-24	5-01-24
Stick in Indicator strip	GNE	GNE	GNE	GNE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5.40pm	4.00pm	4.15pm	4.20pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	11.30.23	11.30.23	11.30.23	11.30.23

User, sterilizer & batch number	VH11TB110	VH11TB110	VH11TB110	VH11TB110
Date of production	16-12-23	18-12-23	18-12-23	19-12-23
Expiry Date	5-01-24	6-01-24	6-01-24	8-01-24
Stick in Indicator strip	GNE	GNE	GNE	GNE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5.45pm	4.30pm	5.45pm	4.40pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	11.30.23	11.30.23	11.30.23	11.30.23

SAL Steri-test BOWIE&DICK TEST PACK



DC
SHNU DENTAL COLLEGE

STEAM

LOT 230205

121°C / 10 min.

134°C / 3,5 min.

cess

3:

Batch Monitoring*

User, sterilizer & batch number	VH105B110	VH105B110	VH105B110	VH105B110
Date of production	19-12-23	20-12-23	20-12-23	21-12-23
Expiry Date	8-01-24	8-01-24	8-01-24	10-01-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:50PM	5:00PM	6:15PM	5:00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10:30pm	10:30pm	10:30pm	10:30pm

User, sterilizer & batch number	VH105B110	VH105B110	VH105B110	VH105B110
Date of production	21-12-23	22-12-23	23-12-23	26-12-23
Expiry Date	10-01-24	10-01-24	10-01-24	10-01-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	6:15PM	4:50PM	5:00PM	5:15PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10:30pm	10:30pm	10:30pm	10:30pm

Sterilization

Bowie-Dick

BDS-t

Batch Moni

User, steriliz
batch num

Date of p


Expiry Date

Stick in Indicat

Program

Temperature
Starting Time

SAL Steri-test BOWIE&DICK TEST PACK



STEAM

LOT 230205

121°C / 10 min.

134°C / 3,5 min.

3.10

-23

24

gph

10:00 PM

Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

User, sterilizer & batch number	VHICJB110	VHICJB110	VHICJB110	VHICJB110
Date of production	30-12-23	31-12-23	2-01-24	3-01-24
Expiry Date	15-01-24	15-01-24	21-01-24	21-01-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garretcycle	Garretcycle	Garretcycle	Garretcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:20 PM	5:30 PM	5:15 PM	12:00 AM
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanthi	K. Jayanthi	K. Jayanthi	K. Jayanthi

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH105B110	VH105B110	VH105B110	VH105B110
Date of production	3-01-24	4-01-24	5-01-24	5-01-24
Expiry Date	21-01-24	21-01-24	21-01-24	21-01-24
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:20 PM	5:30 PM	12:00 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10.3.2024	10.3.2024	10.3.2024	10.3.2024

User, sterilizer & batch number	VH105B110	VH105A110	VH105A110	VH105A110
Date of production	6-01-24	8-01-24	8-01-24	9-01-24
Expiry Date	25-01-24	27-01-24	27-01-24	28-01-24
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:30 PM	1:10 PM	5:20 PM	1:20 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	10.3.2024	10.3.2024	10.3.2024	10.3.2024

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH US A110	VH US B110	VH US A110	VH US B110
Date of production	10-01-24	10-01-24	11-01-24	11-01-24
Expiry Date	29-01-24	29-01-24	30-01-24	30-01-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	12.00 PM	5.40 PM	5.30 PM	5.45 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. Jayanthi	U. Jayanthi	U. Jayanthi	U. Jayanthi

User, sterilizer & batch number	VH US A110	VH US B110	VH US A110	VH US B110
Date of production	12-01-24	12-01-24	13-01-24	18-01-24
Expiry Date	30-01-24	30-01-24	30-01-24	6-02-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	1.00 PM	5.20 PM	5.10 PM	5.30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. Jayanthi	U. Jayanthi	U. Jayanthi	U. Jayanthi

SAL Steri-test BOWIE&DICK TEST PACK

20-01-24 62(B)



Docun

Sterilizer No. _____

Sterilization Proc _____

Bowie-Dick-Simu _____

BDS-test inc _____

STEAM

LOT 230205

121°C / 10 min.


134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH113A110	VH113R110	VH113A110	VH113R110
Date of production	19-01-24	20-01-24	22-01-24	23-01-24
Expiry Date	7-02-24	8-02-24	10-02-24	10-02-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30pm	5:30pm	5:20pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	Kesavanth	Kesavanth	Kesavanth	Kesavanth

User, sterilizer & batch number	VH113A110	VH113A110	VH113R110	VH113R110
Date of production	24-01-24	24-01-24	25-01-24	25-01-24
Expiry Date	12-02-24	12-02-24	13-02-24	13-02-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	1:10pm	5:30pm	1:15pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	Kesavanth	Kesavanth	Kesavanth	Kesavanth

SAL Steri-test BOWIE & DICK TEST PACK



STEAM

LOT 230205

121°C / 10 min.
134°C / 3,5 min.

Sterilizer No. _____
Sterilization Program _____
Bowie-Dick-Simulation _____

BDS-test indicator _____

Batch Monitoring*

User, sterilizer & batch number	VH115 A110	VH115 A110	VH115 R110	VH115 R110
Date of production	27-01-24	27-01-24	29-01-24	29-01-24
Expiry Date	15-02-24	15-02-24	16-02-24	16-02-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	1:20 PM	5:30 PM	1:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. S. J. G. L.	U. S. J. G. L.	U. S. J. G. L.	U. S. J. G. L.

User, sterilizer & batch number	VH115 A110	VH115 A110	VH115 R110	VH115 R110
Date of production	30-01-24	30-01-24	31-01-24	01-02-24
Expiry Date	17-02-24	17-02-24	18-02-24	20-02-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	1:20 PM	5:40 PM	5:20 PM	5:30 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	U. S. J. G. L.	U. S. J. G. L.	U. S. J. G. L.	U. S. J. G. L.

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer N

Sterilization

Bowie-Dick

BDS-t

STEAM

LOT 230205


121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH10TA110		VH10TA110		VH10TR110		VH10TR110	
Date of production	2-02-24		2-02-24		3-02-24		3-02-24	
Expiry Date	21-02-24		21-02-24		22-02-24		22-02-24	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121°C		121°C		121°C		121°C	
Starting Time	1:20PM		5:20PM		1:20PM		5:20PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanth		K. Jayanth		K. Jayanth		K. Jayanth	

User, sterilizer & batch number	VH10TA110		VH10TA110		VH10TR110		VH10TR110	
Date of production	5-02-24		5-02-24		6-02-24		6-02-24	
Expiry Date	24-02-24		24-02-24		25-02-24		25-02-24	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121°C		121°C		121°C		121°C	
Starting Time	5:00PM		6:00PM		1:20PM		5:15PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. Jayanth		K. Jayanth		K. Jayanth		K. Jayanth	

SAL Steri-test BOWIE&DICK TEST PACK



Access

te:

STEAM

LOT 230205

121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH115A110	VH115A110	VH115R110	VH115R110
Date of production	7-02-24	7-02-24	8-02-24	8-02-24
Expiry Date	26-02-24	26-02-24	27-02-24	27-02-24
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:00PM	5:30PM	4:10PM	5:45PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

User, sterilizer & batch number	VH115A110	VH115A110	VH115R110	VH115R110
Date of production	9-02-24	9-02-24	10-02-24	10-02-24
Expiry Date	28-02-24	28-02-24	29-02-24	29-02-24
Stick in Indicator strip				
Program	Garment Cycle	Garment Cycle	Garment Cycle	Garment Cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:15PM	5:30PM	4:20PM	5:45PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer N
Sterilization
Bowie-Dick



BD
[]

STEAM

LOT 230205

121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH11TA110	VH11TA210	VH11TB110	VH11TB210
Date of production	12-02-24	12-02-24	13-02-24	13-02-24
Expiry Date	2-03-24	2-03-24	3-03-24	3-03-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	4:10 PM	5:40 PM	4:20 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	[Signature]	[Signature]	[Signature]	[Signature]

User, sterilizer & batch number	VH11TA110	VH11TA210	VH11TB110	VH11TB210
Date of production	14-02-24	14-02-24	15-02-24	16-02-24
Expiry Date	4-03-24	4-03-24	5-03-24	6-03-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	4:10 PM	5:40 PM	5:30 PM	5:40 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	[Signature]	[Signature]	[Signature]	[Signature]

92-02-2164(B)

SAL Steri-test BOWIE&DICK TEST PACK



Sterilize

Sterilization

Bowie-Dick

BD

STEAM

LOT 230205


121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH KJ B1 06		VH KJ B1 06	
Date of production	17-02-24		17-02-24	
Expiry Date	7-03-24		12-03-2024	
Stick in Indicator strip	GKE		GKE	
Program	Garment cycle		Garment cycle	
Temperature	121°C		121°C	
Starting Time	4:10 PM		5:30 PM	
Test OK?	☒ Yes ☐ No		☒ Yes ☐ No	
Signature	K. S. J. J.		K. S. J. J.	

User, sterilizer & batch number	VH KJ A 1 06		VH KJ A 2 06	
Date of production	20-02-2024		21-02-2024	
Expiry Date	12-03-2024		12-03-2024	
Stick in Indicator strip	GKE		GKE	
Program	Garment cycle		Garment cycle	
Temperature	121°C		121°C	
Starting Time	4:00 PM		5:30 PM	
Test OK?	☒ Yes ☐ No		☒ Yes ☐ No	
Signature	K. S. J. J.		K. S. J. J.	

SAL Steri-test BOWIE&DICK TEST PACK



De
Sterilizer No.
Sterilization
Bowie-Dick-

BDS-test	STEAM	LOT 230205	121°C / 10 min. 134°C / 3,5 min.
☐ YES		☐ NO	


Batch Monitoring*

User, sterilizer & batch number	VH1C1A106	VH1C5B106	VH1C1A110	VH1C5A110
Date of production	22-02-24	23-02-24	24-02-24	24-02-24
Expiry Date	12-03-24	13-03-24	14-02-24	14-03-24
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	6:00pm	5:00pm	4:00pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

User, sterilizer & batch number	VH1C5B110	VH1C5B110	VH1C1A110	VH1C5A110
Date of production	26-02-24	26-02-24	27-02-24	27-02-24
Expiry Date	14-03-24	14-03-24	15-02-24	15-03-24
Stick in Indicator strip				
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:00pm	5:40pm	4:00pm	5:40pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

SAL Steri-test BOWIE&DICK TEST PACK

28-02-24




D
Sterilizer No
Sterilization
Bowie-Dick

BDS-te	STEAM	LOT 230205	121°C / 10 min. 134°C / 3,5 min.
--------	--------------	------------	-------------------------------------

Batch Monitoring*

User, sterilizer & batch number	WHITR100		WHITR100		WHITR100		WHITR100	
Date of production	28-02-24		28-02-24		29-02-24		29-02-24	
Expiry Date	16-03-24		16-03-24		17-03-24		17-03-24	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	4:10 PM		4:40 PM		4:15 PM		4:45 PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. S. Gupta		K. S. Gupta		K. S. Gupta		K. S. Gupta	

User, sterilizer & batch number	WHITR100		WHITR100		WHITR100		WHITR100	
Date of production	1-03-24		1-03-24		2-03-24		2-03-24	
Expiry Date	20-03-24		20-03-24		21-03-24		21-03-24	
Stick in Indicator strip	GKE		GKE		GKE		GKE	
Program	Garment cycle		Garment cycle		Garment cycle		Garment cycle	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	4:00 PM		5:30 PM		4:10 PM		5:40 PM	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	K. S. Gupta		K. S. Gupta		K. S. Gupta		K. S. Gupta	



Sterili:

Sterili

Bowie

STEAM

LOT 230937

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHUTR 110	VHUTR 110	VHUTR 110	VHUTR 110
Date of production	4-03-24	4-03-24	03-03-2024	03-03-2024
Expiry Date	23-03-24	23-03-24	20-03-24	20-03-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentoph	Garmentoph	Garmentoph	Garmentoph
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:00pm	5:40pm	4:00pm	5:50pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Singh	K. Singh	K. Singh	K. Singh

User, sterilizer & batch number	VHUTR 110	VHUTR 110	VHUTR 110	VHUTR 110
Date of production	6-03-24	7-03-24	9-03-24	11-03-24
Expiry Date	25-03-24	26-03-24	28-03-24	29-03-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentoph	Garmentoph	Garmentoph	Garmentoph
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:30pm	4:55pm	5:30pm	4:55pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	K. Singh	K. Singh	K. Singh	K. Singh

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH115A110	VH115A110	VH115A110	VH115A110
Date of production	12-03-24	12-03-24	13-03-24	13-03-24
Expiry Date	31-03-24	31-03-24	1-04-24	1-04-24
Stick in Indicator strip				
Program	Fast cycle	Fast cycle	Fast cycle	Fast cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:00 PM	5:45 PM	4:10 PM	5:50 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

User, sterilizer & batch number	VH115A110	VH115A110	VH115A110	VH115A110
Date of production	14-03-24	14-03-24	15-03-24	15-03-24
Expiry Date	2-04-24	2-04-24	3-04-24	3-04-24
Stick in Indicator strip				
Program	Fast cycle	Fast cycle	Fast cycle	Fast cycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:00 PM	5:30 PM	4:10 PM	5:45 PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer

Sterilizat

Bowie-D

BD



STEAM

LOT 230937

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH103A10	VH103A10	VH103A10	VH103A10
Date of production	16-03-24	17-03-24	18-03-24	19-03-24
Expiry Date	14-04-24	15-04-24	16-04-24	17-04-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:00PM	4:00PM	5:45PM	4:00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	103-24	103-24	103-24	103-24

User, sterilizer & batch number	VH103A10	VH103A10	VH103A10	VH103A10
Date of production	19-03-24	20-03-24	20-03-24	21-03-24
Expiry Date	7-04-24	08-04-24	8-04-24	9-04-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:45PM	4:00PM	5:45PM	4:00PM
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	103-24	103-24	103-24	103-24

SAL Steri-test BOWIE & DICK TEST PACK



Sterilizer

Sterilizat

Bowie-Di

BD

STEAM

LOT 230937

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH115 A110		VH115 A110		VH115 A110		VH115 A110	
Date of production	21-03-24		22-03-24		22-03-24		23-03-24	
Expiry Date	9-04-24		10-04-24		10-04-24		11-04-24	
Stick in Indicator strip	GNE		GNE		GNE		GNE	
Program	Garment		Garment		Garment		Garment	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	5:50pm		6:10pm		5:45pm		6:00pm	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	[Signature]		[Signature]		[Signature]		[Signature]	

User, sterilizer & batch number	VH115 A110		VH115 A110		VH115 A110		[Redacted]	
Date of production	23-03-24		25-03-24		25-03-24		26-03-24	
Expiry Date	11-04-24		13-04-24		13-04-24		[Redacted]	
Stick in Indicator strip	GNE		GNE		GNE		GNE	
Program	Garment		Garment		Garment		Garment	
Temperature	121 °C		121 °C		121 °C		121 °C	
Starting Time	5:40pm		6:10pm		5:40pm		5:50pm	
Test OK?	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No	☒ Yes	☐ No
Signature	[Signature]		[Signature]		[Signature]		[Signature]	

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Documentation Sheet to Monitor CSSD Sterilization Process

Sterilizer No. _____

Date: _____

Sterilization Process: Steam

Bowie-Dick-Simulation Test (BDS)

BDS-test indicator strip	Test OK?	Responsible User
	☐ Yes ☐ No	

Batch Monitoring*

User, sterilizer & batch number	VH/11TA/11	VH/11TA/11	VH/11TA/11	VH/11TA/11
Date of production	27-03-24	29-03-24	30-03-24	01-4-24
Expiry Date	16-04-24	17-04-24	18-04-24	20-4-24
Stick in Indicator strip				
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:40pm	5:40pm	5:40pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

User, sterilizer & batch number	VH/11TA/11	VH/11TA/11	VH/11TA/11	VH/11TA/11
Date of production	2-4-24	3-4-24	4-4-24	5-4-24
Expiry Date	21-4-24	22-4-24	23-4-24	24-4-24
Stick in Indicator strip				
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:40pm	5:30pm	5:40pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature				

SAL Steri-test BOWIE & DICK TEST PACK



Sterilizer

Sterilization

Bowie-Dick

BD

STEAM

LOT 230937

121°C / 10 min.


134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH103 A110	VH103 A110	VH103 A110	VH103 A110
Date of production	6-04-24	8-04-24	10-04-24	11-04-24
Expiry Date	25-04-24	25-04-24	25-04-24	26-04-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:30pm	5:40pm	5:50pm	5:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	103-110	103-110	103-110	103-110

User, sterilizer & batch number	VH103 A110	VH103 A110	VH103 A110	VH103 A110
Date of production	12-04-24	13-04-24	15-04-24	16-04-24
Expiry Date	27-04-24	28-04-24	29-04-24	30-04-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garment cycle	Garment cycle	Garment cycle	Garment cycle
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	5:40pm	5:45pm	5:50pm	5:40pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	103-110	103-110	103-110	103-110

SAL Steri-test BOWIE & DICK TEST



Do
Sterilizer No.
Sterilization
Bowie-Dick-S

BDS-test

STEAM

LOT 230937

121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VHUS A10	VHUS A10	VHUS A10	VHUS A10
Date of production	17-04-24	18-04-24	19-04-24	19-04-24
Expiry Date	6-05-24	7-05-24	8-05-24	8-05-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	5:30pm	5:40pm	5:45pm	6:45pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	W. J. J. J.	W. J. J. J.	W. J. J. J.	W. J. J. J.

User, sterilizer & batch number	VHUS R10	VHUS A10	VHUS A10	VHUS R10
Date of production	20-04-24	20-04-24	21-04-24	22-04-24
Expiry Date	9-05-24	9-05-24	11-05-24	11-05-24
Stick in Indicator strip	GKE	GKE	GKE	GKE
Program	Garmentcycle	Garmentcycle	Garmentcycle	Garmentcycle
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:40pm	5:40pm	4:50pm	5:40pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	W. J. J. J.	W. J. J. J.	W. J. J. J.	W. J. J. J.

SAL Steri-test BOWIE & DICK TEST PACK



Sterilizer

Sterilization

Bowie-Dick

BDS

STEAM

LOT 230937

121°C / 10 min.

134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH115A110	VH115B110	VH115A110	VH115B110
Date of production	23-04-24	23-04-24	24-04-24	24-04-24
Expiry Date	12-05-24	12-05-24	13-05-24	13-05-24
Stick in Indicator strip				
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:30pm	5:40pm	4:40pm	5:50pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	W. S. J. J. J.	W. S. J. J. J.	W. S. J. J. J.	W. S. J. J. J.

User, sterilizer & batch number	VH115A110	VH115B110	VH115A110	VH115B110
Date of production	25-04-24	25-04-24	26-04-24	26-04-24
Expiry Date	14-05-24	14-05-24	15-05-24	15-05-24
Stick in Indicator strip				
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121°C	121°C	121°C	121°C
Starting Time	4:30pm	5:40pm	4:40pm	5:50pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	W. S. J. J. J.	W. S. J. J. J.	W. S. J. J. J.	W. S. J. J. J.

SAL Steri-test BOWIE&DICK TEST PACK



Sterilizer No. _____

Sterilization P _____

Bowie-Dick-Si _____

BDS-test

STEAM

LOT 230937

121°C / 10 min.
134°C / 3,5 min.

Batch Monitoring*

User, sterilizer & batch number	VH11TA110	VH11TR110	VH11SA110	VH11TR110
Date of production	27-04-24	27-04-24	28-04-24	28-04-24
Expiry Date	16-05-24	16-05-24	16-05-24	16-05-24
Stick in Indicator strip				
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	4:40pm	5:30pm	4:40pm	5:40pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	<i>U. Sengupta</i>	<i>U. Sengupta</i>	<i>U. Sengupta</i>	<i>U. Sengupta</i>

User, sterilizer & batch number	VH11SA110	VH11TR110	VH11SA110	VH11TR110
Date of production	30-04-24	30-04-24	01-05-24	01-05-24
Expiry Date	18-05-24	18-05-24	20-05-24	20-05-24
Stick in Indicator strip				
Program	Garmentech	Garmentech	Garmentech	Garmentech
Temperature	121 °C	121 °C	121 °C	121 °C
Starting Time	4:30pm	5:40pm	4:00pm	2:30pm
Test OK?	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No	☒ Yes ☐ No
Signature	<i>U. Sengupta</i>	<i>U. Sengupta</i>	<i>U. Sengupta</i>	<i>U. Sengupta</i>