

Dental Equipment Compliant Form

Date: 3/5/2023

Name of the Department: OMFS OT

Equipment & Sl. No.: Dental Chair

Compliant Details: ward chair Repair.

P. S. Kumar

P. Shanmugam
Signature of the HOD / Incharge

Service Report: Power Transformer Fuse Burned Change
to New Fuse Tested ok working properly

P. Shanmugam
Signature of the Technician

Remarks by Staff in Charge: Working properly

P. S. Kumar

P. Shanmugam
Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 12/5/23

Name of the Department: UG-2 Comprehensive

Equipment & Sl. No.: Kettle KSD686-CF.

Compliant Details: kettle not working

Y. Anandhan


Signature of the HOD / Incharge

Y. Anandhan

Service Report: check the Complaint charger


(power) switch Complaint Servicing done

Tested OK working properly

Y. An


Signature of the Technician

Remarks by Staff in Charge: Kettle is working


Signature of the HOD / Incharge

Y. Anandhan

Dental Equipment Compliant Form

Date: 25/5/2023

Name of the Department: Oral & Maxillofacial Surgery

Equipment & SI. No.: Suction motor S.no 170618925

Compliant Details: Vanlago Suction not working in OMFS PG

Minor OT

B. Priskelle



Signature of the HOD / Incharge

Service Report: Motor stroked Repair purpus send to
Company through store After 21/07/23 Repair
check up done tested OK working properly

M. Sathya



Signature of the Technician

Remarks by Staff in Charge:

Working properly

B. Priskelle

Ushir 21/7/23
Signature of the HOD / Incharge

to

Dental Equipment Compliant Form

Date: 08.06.2023

Name of the Department: Oral & Maxillo facial surgery

Equipment & Sl. No.: Chair-2 Light (minox or).

Compliant Details: chair Light was not working

U. Sri R.

Signature of the HOD / Incharge

Service Report: Check the complaint power cable
Loose contact Re Connected and
Tested OK working properly

M. Lethis

M. Lethis

Signature of the Technician

Remarks by Staff in Charge:

Now it is working

U. Sri R.

Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 12/06/23

Name of the Department: UG - 08

Equipment & Sl. No.: Lab Micro motor 15084568

Compliant Details: Micro motor - Not working

Y. Annarane

Dr. Anil Kumar
Signature of the HOD / Incharge

Service Report: m-4 H/p Carbon Brushes Damaged
Change the New Carbon Brushes Tested & OK
Working properly

T. M. Sathya

[Signature]
Signature of the Technician

Remarks by Staff in Charge: Verified
Working properly

Y. Annarane

[Signature]
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 12/6/23

Name of the Department: Perio PG

Equipment & Sl. No.: P5 scales S.NO-41113046

Compliant Details: Scaler not working (water spraying not working)

T. Manikyan
12/6/23

[Signature]
Signature of the HOD / Incharge

Service Report: check the complaint water filter
Servicing & water tube servicing done Tested
OK working properly

M. Sathya

[Signature]
Signature of the Technician

Remarks by Staff in Charge: working properly

[Signature]
Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 12/6/23

Name of the Department: pedodontics and preventive dentistry

Equipment & Sl. No.: spot welders # 2 N.S.

Compliant Details: spot welders not working.

P. Harithamma.

QIN
12/6/23

Signature of the HOD / Incharge

Service Report: check the Complai Both units
are servicing done Tested ok working
propurelley

K. S

K. S

Signature of the Technician

Remarks by Staff in Charge: Working Properly

P. Harithamma.

QIN
8/12/23

Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 19.6.23

Name of the Department: Prosthodontics - Pg Clinic

Equipment & Sl. No: Dental chair no: 8 & no: 10

Compliant Details: water leakage from airtor
trolley & airtor valves

Prasanna
19.6.23



Signature of the HOD / Incharge

Service Report: check the complaint both dental chair w.r valves
are fixed. so new water retentive valves replaced done.

H. Bithi



Signature of the Technician

Remarks by Staff in Charge: complaint rectified and
now the chair water leakage stopped.
Chairs are working.

Prasanna
20.6.23



Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 21/06/2023

Name of the Department: UG-4

Equipment & SI. No.: Chairs - 14, 22

Compliant Details: Chair 14, 22 - scaler and aerator not
working

N. Mangatayam.


Signature of the HOD / Incharge

Service Report: Water Retractive valve & Spray valve
servicing done Tested OK working properly

M. Sathya


Signature of the Technician

Remarks by Staff in Charge: working properly

N. Mangatayam.


Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 26/06/23

Name of the Department: Ed. & Maxillo-facial Surgery

Equipment & Sl. No.: Ventay mobile section & sl no 170518711

Compliant Details: Ventay mobile section is not
working

M. Esther Rani

U. Shy Baji
Signature of the HOD / Incharge

Service Report: check the complaint 8mup capacitor failed so new
capacitor replaced done after that checked suction working properly.

M. Esther

h.wan
Signature of the Technician

Remarks by Staff in Charge:

working properly

M. Esther Rani

U. Shy Baji
Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 8/7/23

Name of the Department: Paediatric Endodontology

Equipment & Sl. No.: chair no. 781 & 5 (3 chairs)

Compliant Details: water leaking from spittoon

I. Thine

for HOD

Signature of the HOD / Incharge

Service Report: 3 chairs are jolter cap sitting part disconnected so use the jolter cap properly fixed, now working condition

M. Sathya

for HOD

Signature of the Technician

Remarks by Staff in Charge: Now working properly

K. Manikumar

for HOD

Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 21/07/2023

Name of the Department: Oral & Maxillo facial Surgery

Equipment & Sl. No.: Mobile Venky Suction apparatus, S.No: 170818711

Compliant Details:

Not working

B. Pristelle

U. Shiji Raju
Signature of the HOD / Incharge

Service Report: Motor stucked Repair purpus send to
store After Repair (17.08.23) check up done
working properly

M. Sathya

[Signature]
Signature of the Technician

Remarks by Staff in Charge: Working properly

M. E Rami

U. Shiji Raju
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 2/8/23

Name of the Department: Pedodontics & Preventive Dentistry

Equipment & Sl. No.: Dental chair No: 16

Compliant Details: Light not working, Suction is not working properly.

J. Hinegar

CAV
2/8/23

Signature of the HOD / Incharge

Service Report: change the new bulb & High Suction Tube check up done working properly

M. Sathya

K. V. S.

Signature of the Technician

Remarks by Staff in Charge: Light & Suction also working properly.

J. Hinegar

CAV
2/8/23

Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 12-08-2023

Name of the Department: UG-4 Clinic

Equipment & Sl. No.: Micromotor Hand piece SL No: AA.19.024.007

Compliant Details: Micromotor Hand piece not working

G. Haritha

Signature of the HOD / Incharge

Service Report: Bearings Damage So Replace
purpus Send to Company 08.9.23 After
Repair 21.9.23 Tested are working properly

M. Sathya

Signature of the Technician

Remarks by Staff in Charge: Micromotor is working

G. Harithamma

Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 9/9/23

Name of the Department: Conservative Dentistry and Endodontics

Equipment & Sl. No.: Dental Chair - 04

Compliant Details: foot Control - Not working

M. Saranya Sani
M.D.S.

Signature of the HOD / Incharge

Service Report: Check the Complaint Hydraulic Assemble
Damaged not Reported change the New Hydraulic
Assemble After check up done working properly

M. Subhas
Ch. N. V. Durga Rao

Signature of the Technician

Remarks by Staff in Charge: foot Control returned to working

Condition:

M. Saranya Sani

M.D.S.

13/9/23

Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 14/09/2023

Name of the Department: UG-3

Equipment & Sl. No.: Dental chair No: 13

Compliant Details: light compliant

N. Mangatayam

For Rehman
15/09/23

Signature of the HOD / Incharge

Service Report: Bolt Brass Holder Rest formed

Open the Holder & power cable servicing
done working properly

M. Suresh

d. N.V. Durgar Rao

For Raman

Signature of the Technician

Remarks by Staff in Charge: working properly

N. Mangatayam

For Rehman
26/09/23

Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 21/09/2023

Name of the Department: UG-7

Equipment & Sl. No.: Lab Micromotor No. W.A.A. 18041801

Compliant Details: Not working (Lab micromotor) Since yesterday

Signature of the HOD / Incharge

Y. Annamalai:

Service Report: check the complaint servicing
done Tested ok working properly

M. S. S. S.

Signature of the Technician

Remarks by Staff in Charge: Reviewed lab micromotor in working
condition f

Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 23-09-23

Name of the Department: UG-4

Equipment & Sl. No.: Dental Chair No. 11

Compliant Details: Sound from Sound is coming from the
chair

M. Shanmukhi

Signature of the HOD / Incharge

Service Report: Drain Tube, Section L-Bow & Drain Tube
to placement Reducer 3 Spares are Damaged
Replace the new Spares checkup done working

-propelled

M. Subbar

Signature of the Technician

Remarks by Staff in Charge: Working Properly

G. Harthamure

Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 25/9/23

Name of the Department: Prosthetic (pa. clinic)

Equipment & Sl. No.: Dental chair - (10)

Compliant Details: ① Suction is not working properly

② Head rest is not fitting properly

G. N. N. N. N.
25/9/23

Dr. N. N. N.

Signature of the HOD / Incharge

Service Report: Low suction Tube New Replaced &

Servicing done Head Rest Adjusted Tested & working properly

Dr. N. N. N.

Signature of the Technician

Remarks by Staff in Charge: Working properly

G. N. N. N.
25/9/23

Dr. N. N. N.

Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 6/10/23

Name of the Department: Fast track Clinic

Equipment & SI. No.: Aerotax hand Pieces ① 16K17595 ② 16E30137

Compliant Details: Locking complaint

P.v. Lakshmi

P.v. Lakshmi
Signature of the HOD / Incharge

Service Report: check the Complaint Bar checks are
Demaged change to one (1) new Bar
check given & Both H/p services are
working properly

K. Jay

K. Jay
Signature of the Technician

Remarks by Staff in Charge: new Bar check, working properly.

P.v. Lakshmi

P.v. Lakshmi
Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 9/10/23

Name of the Department: UG-Clinic -2

Equipment & Sl. No.: Dental chair no. 2

Compliant Details: Air leakage from chair

N. Mangatayam


Signature of the HOD / Incharge

Service Report: Main Air line tube Damage
(Drain to Spittin Box) Replace the new
P.C Tube working now

M. Sathya


Signature of the Technician

Remarks by Staff in Charge: Working properly

N. Mangatayam


Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 10/10/2023

Name of the Department: oral & maxillo facial surgery

Equipment & Sl. No.: MARTHON Micromotor Hand pieces, No: 3041519

Compliant Details:

Not working

P. S. Srinivasan

U. S. R. R.
Signature of the HOD / Incharge

Service Report: Micromotor power cord. Damaged

Change the New Stright cord changed
Tested & working properly

(At OT Table working purpus change
new Stright cord)

R. V. V.
Signature of the Technician

Remarks by Staff in Charge:

Received in working condition

U. S. R. R.
Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 11/10/23

Name of the Department: pedodontics & preventive dentistry

Equipment & Sl. No.: Chair NO. 5

Compliant Details: light is not working, spitting suction getting
sounds

G.D.R. Deepika

11/10/23
Signature of the HOD / Incharge

Service Report: Light P.C.B Component Burned
New Component Re Soldering done
Tested ok working properly

K. S. —

18/10/23
Signature of the Technician

Remarks by Staff in Charge: Working properly

G.D.R. Deepika

23/10/23
Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 1/11/2023

Name of the Department: Oral & maxillo facial Surgery

Equipment & Sl. No.: Stright Hand peice FA 21120293

Compliant Details:

Not working properly (struck).

M. E. Lami

U. S. Ravi
Signature of the HOD / Incharge

Service Report: Both 4/p are oil servicing

done now working properly

K. S. —

J. W. A.
Signature of the Technician

Remarks by Staff in Charge:

working properly

M. E. Lami

U. S. Ravi
Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 10/11/23

Name of the Department: ug 4

Equipment & Sl. No.: Straight Handpiece

Compliant Details: straight Handpiece is not working

DH100068

P. Div.
Signature of the HOD / Incharge

Y. Annamalai

Service Report: Check the complaint oil servicing & clean the rest parts after check up done working properly

Sivan
Signature of the Technician

K. S.

Remarks by Staff in Charge: Handpiece (Straight) is working. checked.

P. Div.
Signature of the HOD / Incharge



VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 15/11/23

Name of the Department: UG-3

Equipment & Sl. No.: Dental Chair No -13

Compliant Details: section complaint

G. Narthana

Signature of the HOD / Incharge

Service Report: Section Bush O-Ring Damage
Change the new o-ring check up due
working properly

M. Sathya

Signature of the Technician

Remarks by Staff in Charge: Working properly

Signature of the HOD / Incharge

Dental Equipment Compliant Form

Date: 4/12/2023

Name of the Department: oral and maxillo facial surgery (P.G)

Equipment & Sl. No.: 170618892 mobile suction apparatus

Compliant Details: Motor was not working

B. Priskilla.

P. Arav
Signature of the HOD / Incharge

Service Report: check the complaint capacitor
burned change the new capacitor
up done working properly

M. Sathya

M. Sathya
Signature of the Technician

Remarks by Staff in Charge:

working properly

M. Sathya

P. Arav
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 30/1/24

Name of the Department: Periodontics

Equipment & Sl. No.: Dental chair No. 6

Compliant Details: Dental chair light is not working

K. Manikumar
sister

for [Signature]
Signature of the HOD / Incharge

Service Report: complaint checked, p.c.b wire shouldering
blown wire shouldering blown wire shouldering done.
Then the light will work.

Mr. Sathya

[Signature]
Signature of the Technician

Remarks by Staff in Charge: Working properly

[Signature]

for [Signature]
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 30/1/24

Name of the Department: Periodontics

Equipment & Sl. No.: Dental chair No. 6

Compliant Details: Dental chair light is not working

K. Manikumar
sister

for [Signature]
Signature of the HOD / Incharge

Service Report: complaint checked, p.c.b wire shouldering
blown wire shouldering blown wire shouldering done.
Then the light will work.

Mr. Sathya

[Signature]
Signature of the Technician

Remarks by Staff in Charge: Working properly

[Signature]

for [Signature]
Signature of the HOD / Incharge

Dental Equipment Compliant Form


Date: 19/2/2024

Name of the Department: Ceramic Lab.

Equipment & Sl. No. ^{Lab} ~~480~~ Micromotor handpiece - 18041806; 18105682

Compliant Details: 18041806 - Not working


18105682 - Sound coming



Signature of the HOD / Incharge


Service Report: 1) 18041806 power Card Damaged
Charge Card 2) 682 is Bearing Complaint
Charge Bearing Tested & Working
— Condition

K. S —



Signature of the Technician

Remarks by Staff in Charge: Old bearings are used. Use till problem arises, then new bearings will be provided. Presently bearings are put for indent.



Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 19/12/2024

Name of the Department: Ceramic Lab

Equipment & Sl. No. Lab Micrometer base unit - 601186

Compliant Details: Not working

Signature of the HOD / Incharge

Service Report: Power Switch Loose Contact so

Switch Cables Re Soldering & general

Servicing done working properly

M. Dutt

Signature of the Technician

Remarks by Staff in Charge: Working properly

Signature of the HOD / Incharge

Dental Equipment Compliant Form


Date: 19/2/2024

Name of the Department: Ceramic Lab.

Equipment & Sl. No. ^{Lab} ~~480~~ Micromotor handpiece - 18041806; 18105682

Compliant Details: 18041806 - Not working


18105682 - Sound coming



Signature of the HOD / Incharge


Service Report: 1) 18041806 power Card Damaged
Charge Card 2) 682 is Bearing Complaint
Charge Bearing Tested & Working
— Condition

K. S —



Signature of the Technician

Remarks by Staff in Charge: Old bearings are used. Use till problem arises, then new bearings will be provided. Presently bearings are put for indent.



Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 19.3.24

Name of the Department: Prosthodontics Pg clinic

Equipment & Sl. No.: Dental chair no: 15

Compliant Details: Dental chair foot control not working (up & down)

N. Sri Ganga
Pg Ganga N

Signature of the HOD / Incharge

Service Report: check the complaint Joystick Damage
change the old Joystick Assembly now
working properly

Signature of the Technician

K. J —

Remarks by Staff in Charge: Working properly

N. Sri Ganga

Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 28/03/24

Name of the Department: Comprehensive UG-7 Clinic

Equipment & Sl. No.: Chair No-1 (Dental)

Compliant Details: Light not working

Name
N. Ramesh

P. Dixi
Signature of the HOD / Incharge

Service Report: Light Switch Damage so
change the new Toggle switch check up
done working properly

M. Dattaraj

Signature of the Technician

Remarks by Staff in Charge: In working condition

P. Dixi
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 30/03/24

Name of the Department: Comprehensive UG-2 Clinic

Equipment & Sl. No.: Chair No. 11 (Dental)

Compliant Details: Suction not working

Handwritten signature
N. Ramesh

Handwritten signature
Signature of the HOD / Incharge

Service Report: check the Complaint suction P.C.B
Burned so change the new P.C.B Claged
Tested ok working properly

Handwritten signature
M. Senthil

Handwritten signature
Signature of the Technician

Remarks by Staff in Charge: In working condition

Handwritten signature
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM
Dental Equipment Compliant Form

Date: 18.04.24

Name of the Department: Vg-4

Name of the Equipment: Dental Chair

Serial Number: 20 - chair No.

Compliant Details: Scalar not working

Outlet of water through scalar handpiece slot
is not working.

N. Mangatayaru.

Sister Signature

P. D. S.

Signature of the HOD / Incharge

Service Report: Check the complaint scalar foot control
power cable damaged change the new
cable now working properly

a. Durga Rao

Technician Signature

18/04/24

Incharge of the Technician

Remarks by Staff in Charge: scalar problem is solved

Now it is working properly.

Water output is in working condition

N. Mangatayaru.

Sister Signature

P. D. S.

Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 17/4/24

Name of the Department: D.M & P.C.P

Equipment & Sl. No. 11080652 Lab aileromotor Hfp

Compliant Details: Hand piece repair

P. D. [Signature]

[Signature]
Signature of the HOD / Incharge

Service Report: Check the Complaint Re fitting
& Servicing done Tested on working
propeller

M. Juthu

[Signature]
Signature of the Technician

Remarks by Staff in Charge: Working Propeller

PSVD 134
54

[Signature]
Signature of the HOD / Incharge



VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM
Dental Equipment Compliant Form

Date: 18.04.24

Name of the Department: Vg-4

Name of the Equipment: Dental Chair

Serial Number: 20 - chair No.

Compliant Details: Scalar not working

Outlet of water through scalar handpiece slot
is not working.

N. Mangatayaru.

Sister Signature

P. D. S.

Signature of the HOD / Incharge

Service Report: Check the complaint scalar foot control
power cable damaged change the new
cable now working properly

a. Durga Rao

Technician Signature

18/04/24

Incharge of the Technician

Remarks by Staff in Charge: scalar problem is solved

Now it is working properly.

Water output is in working condition

N. Mangatayaru.

Sister Signature

P. D. S.

Signature of the HOD / Incharge