VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF - ~~ACRYLIC LAB.~~ (PROSTHO)

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 30.09.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- Not Working.

ITEMS : LAB MICROMOTOR (BASE ONLY)

MAKE : MARATHON-4

SUPPLIER : T.D.S

NO OF UNITS : ONE

UNIT SL.No : 1944619

COMPLAINT : Power suppely cemplaint
Not working.

working condition
Received by
D.Vvvarasimhakar

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully
K.S. Varma
(K.S. VARMA)

Umesh 01/10/20

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF CERAMIC LAB

From,
The MAINTENANCE DEPT,
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 16.09.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:-

ITEMS : VITA VACUMAT 40T

MAKE : —

SUPPLIER : NOVA DENTAL PRODUCTS

NO OF UNITS : ONE

UNIT SL.No : NR: 820040328

COMPLAINT : Lift Block

working Condition
Received
Good

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

DATE OF HANDOVER TO STORES:

Received
Raghava Rao
16/9/20

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF ACRYLICK CAR (PROSTHO

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 05.10.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- MOTOR BURNING

ITEMS : MODELTRIMMER C-49

MAKE : CONFIDENT

SUPPLIER : " "

NO OF UNITS : ONE

UNIT SL.No : 108

COMPLAINT : Motor winding & Cooling fan (plastic) Burned Rewinding Purpus Send to Local Workshop Through Store

working condition 12/11/2020

Received in working condition 12.11.20

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF CONS P.G

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 13.11.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- MOTOR BEARIENG COMPLENT

ITEMS : DURR VS 1200S

MAKE : DURR DENTAL

Working Condition
04/12/2020
N. Bala.
04/12/2020

SUPPLIER :

NO OF UNITS : ONE

UNIT SL.No : C 2350 25002 (REF 7138-10.VS 1200S)

COMPLAINT : Bearieng Damage

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF CONS P.G.

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 13.11.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- CONTROL BOX NOT WORKING

ITEMS : CHAIR CONTROL BOX

MAKE : CONFIDENT

SUPPLIER : CONFIDENT

NO OF UNITS : ONE

UNIT SL.No : P.C.B-E.C.C-PL704

COMPLAINT : Chair Up & Down Not working.

Working Condition
Received by
N. Bala.
26/12/2020

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully
(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF UG-2

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 17-12-2020

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- Hand piece Not working.

ITEMS : clinical micromotor.

MAKE : BEING

SUPPLIER : I.D.S

NO OF UNITS : ONE

UNIT SL.No : 16F160105

COMPLAINT : Coupledge Damage

working OK
Received
N. Mangalayadu

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully
(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF PERIO (P.G)

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 17.12.20

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- HAND PIECES ARE NOT WORKING

ITEMS : STRIGHT & CONTRANGLE HAND PIECES

MAKE : MARATHON & NSK

SUPPLIER : IDS

NO OF UNITS : 02

UNIT SL.No : FA1F010465, LE1210323 not Repairbel

COMPLAINT : Beariengs Demage & colredge Demage

marthan Handpice
Repair done
Received
Recived o'kyem
K. Mani kyem
11/1/21

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF PERIO PG

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 04.01.2021

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- SCALAR UNIT NOT WORKING

ITEMS : SCALAR UNIT ONLY

MAKE : ~~EATEC~~ (P5) SATELEC

SUPPLIER : —

NO OF UNITS : ONE

UNIT SL.No : 4049102028

COMPLAINT : VIBRATION & WATER PROPERELLEY NOT - WORKING

Body & Holder, Cable change the New Spare parts so S.No changed.
New No:- 41336 2006

Recived working Condition
K. Manikyam.
21/1/21

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully
(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF UG. 3

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 11.01.21

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- P.C.B PROPERELLEY NOT WORKING

ITEMS : M.M. P.C.B

MAKE : CONFIDENT

SUPPLIER : "

NO OF UNITS : ONE

UNIT SL.No : DCS-HM-04-121045

COMPLAINT : P.C.B Properlley not working.
(Power out put Complaint) R-37 COMPLAIN

Working Condition
N-Ratnalam
04/02/21

Ch.NO 23

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

Unused

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF CONS PG

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 11.01.24

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- chair life, m.m. & Scaler control P.C.B - Propervalley Not working.

ITEMS : P.C.B

MAKE : CONFIDENT

SUPPLIER :

NO OF UNITS : ONE

UNIT SL.No : DCS-4-160914 (DCS-15B-DIGI 1611409)

COMPLAINT : P.C.B Propervalley not working.
(Power cut pala complaint)

Working Condition
04/02/24
N. Bala.
Ch. NO - 18

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

Unit

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF UG - 8

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 24/02/21

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- SCALAR P.C.B'S NOT WORKING

ITEMS : SCALAR ~~P.C.B~~ & mm P.C.B

MAKE : CONFIDENT

SUPPLIER : CONFIDENT

NO OF UNITS : 02

UNIT SL.No : 1) 1611370 2) 1611390

1611370 INDS P.C.B working checked OK
M.M.Tej
16/8/2024

COMPLAINT : Scaler & mm NOT working.

1611390
working checked ok
M. Mangalaja
29/03/2021

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully
(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE & HOSPITAL
VISHNUPUR, BHIMAVARAM - 534202

DEPARTMENT OF PERIODONTICS (P.G)

From,
The MAINTENANCE DEPT.
VISHNU DENTAL COLLEGE::HOSPITAL

Date: 12-03-21

TO,
THE DENTAL STORE,
VISHNU DENTAL COLLEGE:: HOSPITAL
VISHNUPUR,

SUB:- SUCTION MOTOR NOT WORKING

ITEMS : MOBIAL SUCTION UNIT

MAKE : VINKY

SUPPLIER : -

NO OF UNITS : TWO

UNIT SL.No : (1) 1108-12694 (2) 130914644

23/3/21

7/7/21 Recived by K. Mouryam. 7/7/21

COMPLAINT : Two units are motors Complaint

FOR ANY CALARIFICATION: Cel: 9490456448

Yours faithfully

(K.S. VARMA)

DATE OF HANDOVER TO STORES:

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 15.12.20

Name of Department: Prosthodontics - Acrylic Lab

Equipment & Sl.No: Marathon-4 Micromotor Base Units 2

Complaint Details: (1) 700350 (2) 1836618

Not working

Signature of HOD/Incharge

Service Report: Repair privus Send to Cenpius 17/12/20.
After Repair 11/1/21 check up units working properly
Handover to Dept-

Signature of Technician

Remarks by Staff in charge: Received two units in working
Condition

11/1/21

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 17/12/2020

Name of Department: UG-2

Equipment & Sl.No: micro clinical Handpice contrangle / 16.F16.W105

Complaint Details: not working

Signature of HOD/Incharge

Service Report: Repair purpus send to Company 17.12.20 (CATRADGE DAMAGE) After Repair 11/01/21 handover to Dept working condition.

Signature of Technician

Remarks by Staff in charge: Contrangle is working Properly.

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 13/11/20

Name of Department: Conservative Dentistry & Endodontics

Equipment & Sl.No: Chair No:-14 S.No-C235025002

Complaint Details: chair lift up & down not working

N. Bala
Sister.

13/11/20

Signature of HOD/Incharge

Service Report: Control Box Repair pyt pls Send to Confident Service System. Hyd Store After Repair 26/12/20 fix the control b/s check up clear working properly

Signature of Technician

Remarks by Staff in charge: Noted working condition

N. Bala.

Complaint rectified

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 10/12/20

Name of Department: Conservative Dentistry and Endodontics

Equipment & Sl.No: No - 13 Dental chair

Complaint Details: Water not draining properly in the spitoon,
Scaler unit not working.

Complaint rectified

10/12/20

B. V. Sindhuja

Signature of HOD/Incharge

Service Report: Checked the complaint Servicing the Drain Soloined Valve Drain Cuplet (water) soled & Scaler Connections (POWER) Re Soldering done working properly

Signature of Technician

Remarks by Staff in charge: Water drain & Scaler unit were working

B. V. Sindhuja

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 9/3/21

Name of Department: Periodontics (PG)

Equipment & Sl.No: 1108-12694, 1309 14644

Complaint Details: Mabial suction (Vinky) Not working (2ND)

for

Signature of HOD/Incharge

Service Report: Two motors are Cumplent one motor Stocked & one motor is Burning Smell Come So Repair purpose Send to Cumpeny through Store

After Repair 1108-12694 - 23/3/21 working properlly
1309 14644 - 07/07/21

M. Subbarao
Signature of Technician

Remarks by Staff in charge: Recived working

T. Harikyan
7/7/21

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 15.12.20

Name of Department: Prosthodontics - Acrylic lab

Equipment & Sl.No: lab micromotor Hand Piece 19024027

Complaint Details: not working

Signature of HOD/Incharge

Service Report: Repair pr/pus send to Cmpny 17/12/20

After Repair 21/01/21 Tested ok working properlly

Signature of Technician

Remarks by Staff in charge: Received hand piece in working condition on 21/01/21

21/1/21

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 16-2-21

Name of Department: Conservative Dept. 8

Equipment & Sl.No: Chair No: (7) and Chair No: (8)

Complaint Details: Sclay not working

Signature of HOD/Incharge

Service Report: P.C.B's Repair purpose Send to company through Store

P.C.B complaint, D.C.S-15B-Dec 1611370

D.C.S 15-B-16*11390

After Repair 1611370 17/3/21

1611390 27/3/21 Tested OK working properly

M. Sudhame
Signature of Technician

Remarks by Staff in charge: Chairs in working Condition

Signature of HOD/Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Complaint form

Date: 13/3/21

Name of Department: Comprehensive Clinic-2 Clinic

Equipment & Sl.No: Chair No. 21

Complaint Details: Chair light Not working

13/3/21
Signature of HOD/Incharge

Service Report: light P.C.B not working (S.No - 1701044) So this is Papers purpose send to company through store After Repair Replaced P.C.B Tested ok working properlley

M. Sitharam
Signature of Technician

Remarks by Staff in charge: Light is functioning normally.

Signature of HOD/Incharge