VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 2-05-2022

Name of the Department: UG-4

Equipment & Sl. No.: Suction holder

Compliant Details: Suction holder ~~brok~~ dislodged and fallen down.

Harshita

For UG-4 M.S.R

Signature of the HOD / Incharge

Service Report: Suction Holder screws missing new screws Re fitted. working condition

M. Sritharo

Signature of the Technician

Remarks by Staff in Charge: Repair is done and suction is working properly

Harshitta

For UG-4 M.S.R

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 31/5/22

Name of the Department: UG-2

Equipment & Sl. No.: Dental Chair No - 10

Compliant Details: Dental Chair light is not working

D. Uma

Signature of the HOD / Incharge

Service Report: Life Cable Loose Contact servicing done & Suction p.c.B Burned o/d p.c.B Component Replaced Tested ok working properly

K. Subbu

Signature of the Technician

Remarks by Staff in Charge: Working Properly

B. Vijayalakshmi

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 27/06/22

Name of the Department: UG-7 clinic

Equipment & Sl. No.: chair no: (18)

Compliant Details: Light disconnected from chair arm.

Signature of the HOD / Incharge

Service Report: Life Joint Screws Slipped power Cables Soldering disconnected Re Soldering & Screws Refitted tested OK working properlly

K. Sub

Signature of the Technician

Remarks by Staff in Charge: working condition

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 29/07/2022

Name of the Department: Prosthodontics

Equipment & Sl. No.: (06) - Dentalchair

Compliant Details: water leakage from airotor pipeline

Signature of the HOD / Incharge

Service Report: Change the Reetractive Valve
Tested ok working properallay

Signature of the Technician

Remarks by Staff in Charge: complaint Solved
chair is wokking

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 29/8/22

Name of the Department: UG-2

Equipment & Sl. No.: DENTAL CHAIR NO - 13,4

Compliant Details: No water from Air Rotor & chair movements not working (Both chairs)

N. Krupa Bai

Signature of the HOD / Incharge

Service Report: Ch NO 4 Joystick Switch Re soldering. Ch. 13 Control Box power Supply Femal Jack Damage (Lifting Motor Cable) old Box Jack Replaced & Soldering work done working properly

M. Sri Hari

Signature of the Technician

Remarks by Staff in Charge: In working condition

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 27/09/2022

Name of the Department: Pedodontics & Preventive dentistry

Equipment & Sl. No.: 09 (chair no)

Compliant Details: Water leaking from the airotor tube, Scaler not working.

For 28/9/22.

B. Neelima
III MDS

Signature of the HOD / Incharge

Service Report: W.R. Valve Servicing done & Scaler Soldered Stocked Both Complaint Servicing done working properlly

M. Seshu

Signature of the Technician

Remarks by Staff in Charge: Airotor - No water leakage. Scaler working properly.

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 26-9-22

Name of the Department: UG-8

Equipment & Sl. No.: 13, 14, 15 - Dental chair numbers

Compliant Details: Light is not working for the above 3 chairs.

Y. Annavaram

26/9/22
Signature of the HOD / Incharge

Service Report: Ch.No.14 P.C.B Burning UG-2 Switch Replaced Chair P.C.B changed Ep.ch.no.15 new Bulb changed Both chairs are working.

K. S

Signature of the Technician

Remarks by Staff in Charge: Repair done. Both chairs are functioning.

Y. Annavaram

29/09/22
Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 12.10.22

Name of the Department: Prostho Pg clinic

Equipment & Sl. No.: Dental Chair No: 15

Compliant Details: Airotor & micromotor holder broken need to replace.

Signature of the HOD / Incharge

Service Report: Airotor Holder Replaced (Holder Damaged) problem Solved

M. Sathish

Signature of the Technician

Remarks by Staff in Charge: Airotor holder replaced with new one

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 21.10.2022

Name of the Department: UG-4 clinic

Equipment & Sl. No.: Dental chair No: 1

Compliant Details: Low and high suction not working

G. Marthamma

Signature of the HOD / Incharge

Service Report: High & low suction Tubes damaged Changed the new High & low pipes work done

M. Sathya

Signature of the Technician

Remarks by Staff in Charge: Low and high suction are working properly.

G. Marthamma

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 30/11/22

Name of the Department: Fast Track

Equipment & Sl. No.: Dental Chair No (4)

Compliant Details: Dental chair is not moving upwards & downwards

Signature of the HOD / Incharge

Service Report: Check the Complaint UP & Down Lifting Motor Damnege not Repairbel so Change the new Motor Tested ole working properelly

M. Sudhir

Signature of the Technician

Remarks by Staff in Charge: Working properelly

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 24.11.22

Name of the Department: PCP Lab- Prostho

Equipment & Sl. No.: Lab Micro Motor S.No: 11080652

Compliant Details: lab micromotor H.P. not working.
hand overed to technician.

Raised By

Signature of the HOD / Incharge

Service Report: Check the problem Bearing & lock Compleint
Repair parts Send to Company After Come to 15/12/22
Check up done working properlly

Signature of the Technician

Remarks by Staff in Charge: Working properly

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 17.12.22

Name of the Department: Prostho Pg clinic

Equipment & Sl. No.: Dental chair No: 19

Compliant Details: Water not coming through airotor.

Dr. Deepika.K
(Final yr Pg)
K. Deepika.

Signature of the HOD / Incharge

Service Report: Booster Block stroked (nipples) old Block Replaced Tested ok working condition

M Shekhar

Signature of the Technician

Remarks by Staff in Charge: Replaced new block for water

Complaint rectified
K. Deepika

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 15/12/22

Name of the Department: Conservative dentistry and Endodontics

Equipment & Sl. No.: Microscope — 140904055

Compliant Details: Microscope light intensity is not changing and not getting off

Dr. Manohar
II MDS
15/12/22

Signature of the HOD / Incharge

Service Report: Power Regulator soldering disconnected Re soldering done Tested ok working properly

M. Sathish

Signature of the Technician

Remarks by Staff in Charge: Microscope light is working properly with proper changing light intensity

Dr. Manohar
II MDS
30/12/22

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 27-1-23

Name of the Department: UG-3

Equipment & Sl. No.: Dental Chair Num: 04

Compliant Details: 3-way syringe & Scaler not working

G. Keerthana

Signature of the ~~HOD~~ / Incharge

Service Report: New 3-way Syrenge Replaced (old not Reparbel) & Scaler Servicing done tested on working properlley

M. Sekhar

Signature of the Technician

Remarks by Staff in Charge: New 3-way Syringe & Scaler in working condition

Signature of the ~~HOD~~ / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 24/1/23

Name of the Department: Periodontics

Equipment & Sl. No.: chair no- 8

Compliant Details: Light compliant

T. Manikyam.
24/1/23

for [signature]
Signature of the HOD / Incharge

Service Report: Sevnsor P.C.B. Not working change the New P.C.B Set working properly

M. [signature]

[signature]
Signature of the Technician

Remarks by Staff in Charge: working properly

T. Manikyam
24/1/23

for [signature]
Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 21/1/23

Name of the Department: Dept. of Periodontices

Equipment & Sl. No.: Chair number - 2

Compliant Details: Suction holder breakage

T. Manikyam.
21/1/23

Signature of the HOD / Incharge

Service Report: check for complaint suction Holder breakage
this is not repairable so new section holder replaced
After that suction working condition checked ok.

M. Sathish

Signature of the Technician

Remarks by Staff in Charge: in working condition

T. Manikyam.
21/1/23

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 17/2/23

Name of the Department: Department of Periodontics

Equipment & Sl. No.: Chair No: 3, Chair-7

Compliant Details: Suction is not working, Light complaint &

M. Bhanya

Signature of the HOD / Incharge

Service Report: Suction P.C.B Burned change the P.C.B (Godone chaire P.C.B) & lighte cable loose contact servicing does Both Chaires are Tested working properelley

M. Sathwe

Signature of the Technician

Remarks by Staff in Charge: Working properly

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 3/2/23

Name of the Department: UG-2

Equipment & Sl. No.: Dental chair (3)

Compliant Details: up and down Complaint

Ch. Pavani

M. Mangatayaru

Signature of the ~~HOD~~ / Incharge

Service Report: Motor (Mall function) properelley not working not Repair bel so change the new lifting Motor check up done working properelley

M. Sekhar

Signature of the Technician

Remarks by Staff in Charge: working properly

M. Mangatayaru

Signature of the ~~HOD~~ / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 2/2/23

Name of the Department: VG-2

Equipment & Sl. No.: Dental Chairs 5, 6, 9, 10, 22

Compliant Details: for these chairs scalar is not working

B. Gayathri

M. Mangatayaru

Signature of the ~~HOD~~ / Incharge

Service Report: All chairs are check up done soloined value struck & spral valve complaints servicing done check up done workin properley

M-Sethu

Signature of the Technician

Remarks by Staff in Charge: 5 nos chairs are working condition checked O/C.

N. Mangatayaru.

Signature of the ~~HOD~~ / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 4/3/23

Name of the Department: Acrylic lab

Equipment & Sl. No: Hand Micromotor, S.No 700350

Compliant Details: Power connection is ok, but there is no supply from motor to hand peice & also sound is diff from hand peice.

Signature of the HOD / Incharge

Service Report: F&R Switch Burning (Damage) Removed F&R Switch Servicing done Dirrikeeted Connection Given Now working Condition

Signature of the Technician

Remarks by Staff in Charge: → Reverse is not working, rest all is in working condition. Need observation

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 6/3/23

Name of the Department: Pedodontics & preventive dentistry

Equipment & Sl. No.: Dental Chair Number - 16

Compliant Details: Chair complaint, light not working
Suction is also not working with pressure

Faisal Haque

7/3/23

Signature of the HOD / Incharge

Service Report: Light switch Replaced (new)
Ep Suction Complaint Servicing done
now Chair working Condition

M. Sasidhar

Signature of the Technician

Remarks by Staff in Charge: Repair completed
Chair working properly

Faisal Haque

12/4/23.

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

VISHNU
UNIVERSAL LEARNING

Dental Equipment Compliant Form

Date: 06/03/2023

Name of the Department: Pedodontics & Preventive Dentistry

Equipment & Sl. No.: Dental chair No. 02

Compliant Details: Aerotor water leakage ; Scaler not working

K.S.S. Pratyusha

7/3/23

Signature of the HOD / Incharge

Service Report: Airoto & Scalar W-R. Valves & Solenoid Valves are stocked Servicing done Tested ok working properlly

M. Sathish

Signature of the Technician

Remarks by Staff in Charge: Repair done and Chair is working properly

K.S.S. Pratyusha

11/4/23

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 17/04/23

Name of the Department: UG - 2

Equipment & Sl. No.: Dental chair number 2

Compliant Details: Light is not working

Y. Amarnath.

Signature of the HOD / Incharge

Service Report: Check the Complaint Switch Damonage Change the New Toggle Switch. Tested OK working properlly

M. Balaji Varma

Signature of the Technician

Remarks by Staff in Charge: Working properly

Signature of the HOD / Incharge

VISHNU
UNIVERSAL LEARNING

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 26/4/23

Name of the Department: UG-2 Clinic

Equipment & Sl. No.: Dental chair no:-3

Compliant Details: Aerotor water is leaking

Y. Annavaram.

Signature of the HOD / Incharge

Service Report: Check the Complaint Airotor Inner Cupping damage Cheenge the new cupping Tested ok problum Solved

M. Balaji Varma

Signature of the Technician

Remarks by Staff in Charge: Problem rectified

Signature of the HOD / Incharge

VISHNU DENTAL COLLEGE :: BHIMAVARAM

Dental Equipment Compliant Form

Date: 27/4/23

Name of the Department: OG-7

Equipment & Sl. No.: Dental chair no: 19

Compliant Details: Air side air leak (unbearable sound)

Signature of the HOD / Incharge

M. Mangatayaru.

Service Report: Check the Complaint: Basin Bop To Trolley A.V. Supply Tube Damage Change the new B.U. Tube problem Solved

Signature of the Technician

M. Mangatayaru.

Remarks by Staff in Charge: Dr. Anilkumar

working properly

Signature of the HOD / Incharge